Erratum to “Characterization of a 1 mm (DC to 110 GHz) Calibration Kit for VNA”

Chihyun Cho* · Jin-Seob Kang · Joo-Gwang Lee · Hyunji Koo

In the paper entitled “Characterization of a 1 mm (DC to 110 GHz) Calibration Kit for VNA (*Journal of Electromagnetic Engineering and Science*, vol. 19, no. 4, pp. 272–278, 2019)”, a typographical error occurred in Equation (7). The correct equation follows:

$$R_i = \omega L_i = \left(\frac{1}{\pi d} + \frac{1}{\pi D} \right) \sqrt{\frac{\omega \mu_0 \mu_r}{2\sigma}} \quad (7)$$

Group for Electromagnetic Wave Metrology, Division of Physical Metrology, Korea Research Institute of Standards and Science, Daejeon, Korea.

*Corresponding Author: Chihyun Cho (e-mail: chihyun.cho@kriss.re.kr)

This is an Open-Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<http://creativecommons.org/licenses/by-nc/4.0>) which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

© Copyright The Korean Institute of Electromagnetic Engineering and Science. All Rights Reserved.